



CORRECTION


Correction: Comparison of the handheld RETeval ERG system with a routine ERG system in healthy adults and in paediatric patients

Perry Carter · Arabella Gordon-Reid · Fatima Shawkat · Jay E. Self

Published online: 30 October 2020
 © The Author(s), under exclusive licence to The Royal College of Ophthalmologists 2020

Correction to: *Eye*
<https://doi.org/10.1038/s41433-020-01221-2>

The original Article was corrected shortly after publication because Fig. 1 previously had typographical errors. Please find the correct version below and in the original Article.



Reference range		Mixed 'a' wave	Mixed 'b' wave	Dim White 'b' wave	Dim Blue 'b' wave	Photopic 'a' wave	Photopic 'b' wave	Flicker
Amplitude (µV)	Standard	Mean: 16.3	Mean: 46.7	Mean: 35.8	Mean: 31.7	Mean: 10.6	Mean: 24.9	Mean: 15.0
	2.5 th –97.5 th	8.1–29.0	24.0–78.0	19.1–61.0	18.0–56.0	5.0–22.0	11.0–43.9	6.0–29.9
RETeval (trolands)	Mean	14.2	27.1	18.1	25.4	6.2	20.8	23.7
	2.5 th –97.5 th	5.1–23.0	16.0–41.0	10.0–29.0	17.0–39.0	3.0–10.0	11.0–30.0	11.0–36.1
Reference range		Mixed 'a' wave	Mixed 'b' wave	Dim White 'b' wave	Blue 'b' wave	Photopic 'a' wave	Photopic 'b' wave	Flicker
Peak time (ms)	Standard	Mean: 16.5	Mean: 40.1	Mean: 50.0	Mean: 52.8	Mean: 15.1	Mean: 34.2	Mean: 32.1
	2.5 th –97.5 th	12.0–21.0	36.0–46.0	39.1–60.0	44.1–61.0	11.0–21.0	29.1–41.9	27.1–37.0
RETeval (trolands)	Mean	13.1	38.4	63.7	57.4	10.5	29.2	25.5
	2.5 th –97.5 th	10.0–25.0	32.0–49.0	50.0–86.0	48.0–71.0	6.0–14.0	25.0–33.0	24.0–32.0

Fig. 1