




Instrument Name LCMS01 Sample Name LB CVT-01-03 QC

Injection Acq Method QC 5 to 95 MeCN MW 100-800 ESI+- .amx
Name


MWD1B,Sig=220,4 Ref=400,50 PMP1D,Solvent Ratio B



Signal: MWD1B,Sig=220,4 Ref=400,50

RT [min]	Type	Width [min]	Area	Height	Area%	Name
6.740	BV	0.1327	791.6041	209.6925	13.0141	
6.834	VV	0.1644	4896.2749	1333.6354	80.4958	
6.992	VB	0.1558	394.7697	102.7493	6.4901	
		Sum	6082.6487			

MWD1D,Sig=280,4 Ref=400,100 PMP1D,Solvent Ratio B






Signal: MWD1D,Sig=280,4 Ref=400,100

RT [min]	Type	Width [min]	Area	Height	Area%	Name
6.741	BV	0.1135	123.2434	33.1570	11.8858	
6.834	VV	0.1651	835.4167	224.3619	80.5693	



RT [min]	Type	Width [min]	Area	Height	Area%	Name
6.993	VB	0.1414	78.2317	20.9564	7.5448	
		Sum	1036.8917			



Acquisition Method: QC 5 to 95 MeCN MW 100-800 ESI+-.amx

Path: D:\CDSPProjects\CURVETX\Results\01-HIF\LB CVT-01-03 QC.sirslt



Quat. Pump Method

Flow: 0.500 mL/min
Use Solvent Types: Yes
Low Pressure Limit: 0.00 bar
High Pressure Limit: 600.00 bar
Max. Flow Ramp Up: 100.000 mL/min²
Primary Channel: Automatic
Max. Flow Ramp Down: 100.000 mL/min²
Module Display Name: Quat. Pump
Module Type: G7104C
Order: 1

Solvent Composition

Channel	Ch. 1 Solv.	Name 1	Used	Percent (%)
A	100.0 %	Water V.03	Yes	95.00
B	100.0 %	Acetonitrile V.03	Yes	5.00
C			No	
D			No	

Timetable

Time (min)	A (%)	B (%)	C (%)	D (%)	Flow (mL/min)	Pressure (bar)
12.00	0.00	100.00	0.00	0.00	---	---
15.00	0.00	100.00	0.00	0.00	---	---
18.00	95.00	5.00	0.00	0.00	---	---

Stoptime

Stoptime Mode: Time set
Stoptime: 18.00 min