




Instrument Name      LCMS01      Sample Name      LB-CVT-01-10

Injection Acq Method      QC 5 to 95 MeCN MW 100-800 ESI+-.amx  
Name


MWD1B,Sig=220,4 Ref=400,50 PMP1D,Solvent Ratio B



**Signal:** MWD1B,Sig=220,4 Ref=400,50

RT [min]	Type	Width [min]	Area	Height	Area%	Name
5.594	VV	0.2412	3827.6385	979.0012	98.9050	
5.756	VB	0.0896	35.7373	11.3356	0.9234	
5.956	BB	0.1267	6.6383	2.0699	0.1715	
		Sum	3870.0142			

MWD1D,Sig=280,4 Ref=400,100 PMP1D,Solvent Ratio B






**Signal:** MWD1D,Sig=280,4 Ref=400,100

RT [min]	Type	Width [min]	Area	Height	Area%	Name
5.427	BV	0.2175	23.3266	4.7717	4.7803	
5.594	VB	0.2225	464.6445	122.1937	95.2197	



RT [min]	Type	Width [min]	Area	Height	Area%	Name
		Sum	487.9711			



Acquisition Method: QC 5 to 95 MeCN MW 100-800 ESI+-.amx

Path: D:\CDSProjects\CURVETX\Results\LCMS01-2021-08-09 17-46-17+01-00.rslt

## Quat. Pump Method



Flow: 0.500 mL/min  
Use Solvent Types: Yes  
Low Pressure Limit: 0.00 bar  
High Pressure Limit: 600.00 bar  
Max. Flow Ramp Up: 100.000 mL/min<sup>2</sup>  
Primary Channel: Automatic  
Max. Flow Ramp Down: 100.000 mL/min<sup>2</sup>  
Module Display Name: Quat. Pump  
Module Type: G7104C  
Order: 1

#### Solvent Composition

Channel	Ch. 1 Solv.	Name 1	Used	Percent (%)
A	100.0 %	Water V.03	Yes	95.00
B	100.0 %	Acetonitrile V.03	Yes	5.00
C			No	
D			No	

#### Timetable

Time (min)	A (%)	B (%)	C (%)	D (%)	Flow (mL/min)	Pressure (bar)
12.00	0.00	100.00	0.00	0.00	---	---
15.00	0.00	100.00	0.00	0.00	---	---
18.00	95.00	5.00	0.00	0.00	---	---

## Stoptime

Stoptime Mode: Time set  
Stoptime: 18.00 min