


Instrument Name LCMS01 Sample Name LB-CVT-01-19

Injection Acq Method Name QC 5 to 95 MeCN MW 100-800 ESI+-.amx


MWD1B,Sig=220,4 Ref=400,50 PMP1D,Solvent Ratio B



Signal: MWD1B,Sig=220,4 Ref=400,50





RT [min]	Type	Width [min]	Area	Height	Area%	Name
4.125	BV f	0.0533	2.3286	1.4520	0.0880	
4.202	BV	0.1837	341.7949	93.2462	12.9188	
4.315	VV	0.1568	2220.5771	546.9513	83.9308	
4.425	VB	0.1061	81.0247	26.1954	3.0625	
		Sum	2645.7253			

MWD1D,Sig=280,4 Ref=400,100 PMP1D,Solvent Ratio B



Signal: MWD1D,Sig=280,4 Ref=400,100

RT [min]	Type	Width [min]	Area	Height	Area%	Name
4.202	BV	0.1900	37.6829	10.1871	11.7561	
4.315	VB	0.2633	282.8573	67.7868	88.2439	
		Sum	320.5401			





Acquisition Method: QC 5 to 95 MeCN MW 100-800 ESI+-.amx  
Path: D:\CDSPProjects\CURVETX\Results\LCMS01-2021-08-17 17-53-58+01-00.rsl

## Quat. Pump Method

Flow: 0.500 mL/min  
Use Solvent Types: Yes  
Low Pressure Limit: 0.00 bar  
High Pressure Limit: 600.00 bar  
Max. Flow Ramp Up: 100.000 mL/min<sup>2</sup>  
Primary Channel: Automatic  
Max. Flow Ramp Down: 100.000 mL/min<sup>2</sup>  
Module Display Name: Quat. Pump  
Module Type: G7104C  
Order: 1

### Solvent Composition

Channel	Ch. 1 Solv.	Name 1	Used	Percent (%)
A	100.0 %	Water V.03	Yes	95.00
B	100.0 %	Acetonitrile V.03	Yes	5.00
C			No	
D			No	

### Timetable

Time (min)	A (%)	B (%)	C (%)	D (%)	Flow (mL/min)	Pressure (bar)
12.00	0.00	100.00	0.00	0.00	---	---
15.00	0.00	100.00	0.00	0.00	---	---
18.00	95.00	5.00	0.00	0.00	---	---

## Stoptime

Stoptime Mode: Time set  
Stoptime: 18.00 min