




Instrument Name LCMS01 Sample Name LB CVT-01-04 QC

Injection Acq Method QC 5 to 95 MeCN MW 100-800 ESI+-.amx
Name


MWD1B,Sig=220,4 Ref=400,50 PMP1D,Solvent Ratio B



Signal: MWD1B,Sig=220,4 Ref=400,50




RT [min]	Type	Width [min]	Area	Height	Area%	Name
5.342	MM m	0.1356	74.6896	18.2219	3.9593	
5.561	MM m	0.2316	1811.7563	464.7544	96.0407	
		Sum	1886.4459			

MWD1D,Sig=280,4 Ref=400,100 PMP1D,Solvent Ratio B



Signal: MWD1D,Sig=280,4 Ref=400,100

RT [min]	Type	Width [min]	Area	Height	Area%	Name
5.332	BB	0.1894	22.1478	5.2351	9.0764	
5.561	MM m	0.2034	221.8675	56.4940	90.9236	
		Sum	244.0153			



Acquisition Method: QC 5 to 95 MeCN MW 100-800 ESI+-.amx

Path: D:\CDSPROJECTS\CURVETX\Results\01-HIF\LB CVT-01-04 QC.sirslt

Quat. Pump Method

Flow: 0.500 mL/min
Use Solvent Types: Yes
Low Pressure Limit: 0.00 bar
High Pressure Limit: 600.00 bar
Max. Flow Ramp Up: 100.000 mL/min²



Primary Channel: Automatic
Max. Flow Ramp Down: 100.000 mL/min²
Module Display Name: Quat. Pump
Module Type: G7104C
Order: 1

Solvent Composition

Channel	Ch. 1 Solv.	Name	1	Used	Percent (%)
A	100.0 %	Water	V.03	Yes	95.00
B	100.0 %	Acetonitrile	V.03	Yes	5.00
C				No	
D				No	

Timetable

Time (min)	A (%)	B (%)	C (%)	D (%)	Flow (mL/min)	Pressure (bar)
12.00	0.00	100.00	0.00	0.00	---	---
15.00	0.00	100.00	0.00	0.00	---	---
18.00	95.00	5.00	0.00	0.00	---	---

Stoptime

Stoptime Mode: Time set
Stoptime: 18.00 min