




Instrument Name      LCMS01      Sample Name      LB-CVT-01-11

Injection Acq Method      QC 5 to 95 MeCN MW 100-800 ESI+-.amx  
Name


MWD1B,Sig=220,4 Ref=400,50 PMP1D,Solvent Ratio B



**Signal:** MWD1B,Sig=220,4 Ref=400,50

RT [min]	Type	Width [min]	Area	Height	Area%	Name
5.816	MM m	0.0793	58.0886	29.5732	1.9068	
5.909	MV m	0.1354	2887.1060	832.1080	94.7736	
6.009	MM m	0.0754	101.1230	41.7750	3.3195	
		Sum	3046.3176			

MWD1D,Sig=280,4 Ref=400,100 PMP1D,Solvent Ratio B






**Signal:** MWD1D,Sig=280,4 Ref=400,100

RT [min]	Type	Width [min]	Area	Height	Area%	Name
5.738	BV	0.1095	18.1960	5.7322	4.0136	
5.816	VV	0.0699	19.5716	6.3030	4.3170	



RT [min]	Type	Width [min]	Area	Height	Area%	Name
5.909	VV	0.1356	374.9759	105.3051	82.7107	
6.009	VB	0.3517	40.6147	10.2640	8.9586	
		Sum	453.3582			



Acquisition Method: QC 5 to 95 MeCN MW 100-800 ESI+-.amx

Path: D:\CDSProjects\CURVETX\Results\LCMS01-2021-08-09 17-46-17+01-00.rslt



## Quat. Pump Method

Flow: 0.500 mL/min  
Use Solvent Types: Yes  
Low Pressure Limit: 0.00 bar  
High Pressure Limit: 600.00 bar  
Max. Flow Ramp Up: 100.000 mL/min<sup>2</sup>  
Primary Channel: Automatic  
Max. Flow Ramp Down: 100.000 mL/min<sup>2</sup>  
Module Display Name: Quat. Pump  
Module Type: G7104C  
Order: 1

### Solvent Composition

Channel	Ch. 1 Solv.	Name 1	Used	Percent (%)
A	100.0 %	Water V.03	Yes	95.00
B	100.0 %	Acetonitrile V.03	Yes	5.00
C			No	
D			No	

### Timetable

Time (min)	A (%)	B (%)	C (%)	D (%)	Flow (mL/min)	Pressure (bar)
12.00	0.00	100.00	0.00	0.00	---	---
15.00	0.00	100.00	0.00	0.00	---	---
18.00	95.00	5.00	0.00	0.00	---	---

## Stoptime

Stoptime Mode: Time set  
Stoptime: 18.00 min