


Instrument Name LCMS01 Sample Name LB-CVT-01-15

Injection Acq Method Name QC 5 to 95 MeCN MW 100-800 ESI+-.amx


MWD1B,Sig=220,4 Ref=400,50 PMP1D,Solvent Ratio B



Signal: MWD1B,Sig=220,4 Ref=400,50

RT [min]	Type	Width [min]	Area	Height	Area%	Name
5.595	MM n	0.0324	4.6559	4.2491	0.1143	
5.673	MM m	0.1696	3967.7652	1100.6724	97.4488	
5.797	MM m	0.0688	99.2200	36.1122	2.4369	
		Sum	4071.6410			

MWD1D,Sig=280,4 Ref=400,100 PMP1D,Solvent Ratio B






Signal: MWD1D,Sig=280,4 Ref=400,100

RT [min]	Type	Width [min]	Area	Height	Area%	Name
5.673	VV	0.1695	534.2095	145.3769	94.9965	
5.798	VB	0.1785	28.1370	7.0474	5.0035	



RT [min]	Type	Width [min]	Area	Height	Area%	Name
		Sum	562.3465			



Acquisition Method: QC 5 to 95 MeCN MW 100-800 ESI+-.amx

Path: D:\CDSProjects\CURVETX\Results\LCMS01-2021-08-10 17-11-14+01-00.rslt

## Quat. Pump Method



Flow: 0.500 mL/min  
Use Solvent Types: Yes  
Low Pressure Limit: 0.00 bar  
High Pressure Limit: 600.00 bar  
Max. Flow Ramp Up: 100.000 mL/min<sup>2</sup>  
Primary Channel: Automatic  
Max. Flow Ramp Down: 100.000 mL/min<sup>2</sup>  
Module Display Name: Quat. Pump  
Module Type: G7104C  
Order: 1

#### Solvent Composition

Channel	Ch. 1 Solv.	Name 1	Used	Percent (%)
A	100.0 %	Water V.03	Yes	95.00
B	100.0 %	Acetonitrile V.03	Yes	5.00
C			No	
D			No	

#### Timetable

Time (min)	A (%)	B (%)	C (%)	D (%)	Flow (mL/min)	Pressure (bar)
12.00	0.00	100.00	0.00	0.00	---	---
15.00	0.00	100.00	0.00	0.00	---	---
18.00	95.00	5.00	0.00	0.00	---	---

## Stoptime

Stoptime Mode: Time set  
Stoptime: 18.00 min