


Instrument Name LCMS01 Sample Name LB-CVT-01-17

Injection Acq Method Name QC 5 to 95 MeCN MW 100-800 ESI+-.amx


MWD1B,Sig=220,4 Ref=400,50 PMP1D,Solvent Ratio B



Signal: MWD1B,Sig=220,4 Ref=400,50

RT [min]	Type	Width [min]	Area	Height	Area%	Name
6.004	MM m	0.1841	22.2119	13.4273	0.4547	
6.345	BV	0.2462	4698.6589	1243.9234	96.1849	
6.505	VB	0.1405	107.5974	28.5939	2.2026	
6.676	BB	0.2247	56.5592	16.5972	1.1578	
		Sum	4885.0274			




MWD1D,Sig=280,4 Ref=400,100 PMP1D,Solvent Ratio B





Signal: MWD1D,Sig=280,4 Ref=400,100

RT [min]	Type	Width [min]	Area	Height	Area%	Name
6.345	BV	0.2446	596.5187	158.3252	96.2991	
6.505	VB	0.1421	15.6763	4.1276	2.5307	
6.676	BB	0.2333	7.2487	2.1211	1.1702	
		Sum	619.4437			





Acquisition Method: QC 5 to 95 MeCN MW 100-800 ESI+-.amx
Path: D:\CDSPProjects\CURVETX\Results\LCMS01-2021-08-17 17-53-58+01-00.rsl

Quat. Pump Method

Flow: 0.500 mL/min
Use Solvent Types: Yes
Low Pressure Limit: 0.00 bar
High Pressure Limit: 600.00 bar
Max. Flow Ramp Up: 100.000 mL/min²
Primary Channel: Automatic
Max. Flow Ramp Down: 100.000 mL/min²
Module Display Name: Quat. Pump
Module Type: G7104C
Order: 1

Solvent Composition

Channel	Ch. 1 Solv.	Name 1	Used	Percent (%)
A	100.0 %	Water V.03	Yes	95.00
B	100.0 %	Acetonitrile V.03	Yes	5.00
C			No	
D			No	

Timetable

Time (min)	A (%)	B (%)	C (%)	D (%)	Flow (mL/min)	Pressure (bar)
12.00	0.00	100.00	0.00	0.00	---	---
15.00	0.00	100.00	0.00	0.00	---	---
18.00	95.00	5.00	0.00	0.00	---	---

Stoptime

Stoptime Mode: Time set
Stoptime: 18.00 min