




Instrument Name LCMS01 Sample Name LB-CVT-01-22

Injection Acq Method QC 5 to 95 MeCN MW 100-800 ESI+-.amx
Name


MWD1B,Sig=220,4 Ref=400,50 PMP1D,Solvent Ratio B



Signal: MWD1B,Sig=220,4 Ref=400,50

RT [min]	Type	Width [min]	Area	Height	Area%	Name
5.228	BV	0.0968	86.0330	27.4913	1.1723	
5.323	VV	0.1541	1287.2909	361.2168	17.5407	
5.500	VV	0.1864	5664.0611	1543.1539	77.1789	
5.657	VB	0.1628	301.4851	77.7705	4.1081	
		Sum	7338.8700			



MWD1D,Sig=280,4 Ref=400,100 PMP1D,Solvent Ratio B







Signal: MWD1D,Sig=280,4 Ref=400,100

RT [min]	Type	Width [min]	Area	Height	Area%	Name
5.229	BV	0.0976	15.9885	4.9703	1.4361	
5.323	VV	0.1539	188.3822	52.5382	16.9210	
5.500	VV	0.1863	865.4155	236.7599	77.7340	
5.657	VB	0.1622	43.5177	11.1708	3.9089	
		Sum	1113.3039			





Acquisition Method:

QC 5 to 95 MeCN MW 100-800 ESI+-.amx

Path:

D:\CDSProjects\CURVETX\Results\LCMS01-2021-08-17 17-53-58+01-00.rslt

Quat. Pump Method

Flow: 0.500 mL/min
Use Solvent Types: Yes
Low Pressure Limit: 0.00 bar
High Pressure Limit: 600.00 bar
Max. Flow Ramp Up: 100.000 mL/min²
Primary Channel: Automatic
Max. Flow Ramp Down: 100.000 mL/min²
Module Display Name: Quat. Pump
Module Type: G7104C
Order: 1

Solvent Composition

Channel	Ch. 1 Solv.	Name 1	Used	Percent(%)
A	100.0 %	Water V.03	Yes	95.00
B	100.0 %	Acetonitrile V.03	Yes	5.00
C			No	
D			No	

Timetable

Time (min)	A (%)	B (%)	C (%)	D (%)	Flow (mL/min)	Pressure (bar)
12.00	0.00	100.00	0.00	0.00	---	---
15.00	0.00	100.00	0.00	0.00	---	---
18.00	95.00	5.00	0.00	0.00	---	---

Stoptime

Stoptime Mode:

Time set



QC Report

Curve Therapeutics Ltd

Stoptime: 18.00 min