


Instrument Name LCMS01 Sample Name LB-CVT-01-24

Injection Acq Method Name QC 5 to 95 MeCN MW 100-800 ESI+-.amx


MWD1B,Sig=220,4 Ref=400,50 PMP1D,Solvent Ratio B



Signal: MWD1B,Sig=220,4 Ref=400,50

RT [min]	Type	Width [min]	Area	Height	Area%	Name
5.242	BB	0.1289	13.3874	3.5807	0.1612	
5.463	BV f	0.0267	11.6087	11.6210	0.1398	
5.536	BV	0.3322	8171.3974	1878.2332	98.3717	
5.762	VB	0.1604	110.2626	26.3146	1.3274	
		Sum	8306.6561			




MWD1D,Sig=280,4 Ref=400,100 PMP1D,Solvent Ratio B





Signal: MWD1D,Sig=280,4 Ref=400,100

RT [min]	Type	Width [min]	Area	Height	Area%	Name
5.463	BV f	0.0267	1.3573	1.3769	0.1273	
5.536	BV	0.3332	1050.2860	245.9150	98.5437	
5.762	VB	0.1601	14.1638	3.3411	1.3289	
		Sum	1065.8072			





Acquisition Method: QC 5 to 95 MeCN MW 100-800 ESI+-.amx
Path: D:\CDSPProjects\CURVETX\Results\LCMS01-2021-08-17 17-53-58+01-00.rsl

Quat. Pump Method

Flow: 0.500 mL/min
Use Solvent Types: Yes
Low Pressure Limit: 0.00 bar
High Pressure Limit: 600.00 bar
Max. Flow Ramp Up: 100.000 mL/min²
Primary Channel: Automatic
Max. Flow Ramp Down: 100.000 mL/min²
Module Display Name: Quat. Pump
Module Type: G7104C
Order: 1

Solvent Composition

Channel	Ch. 1 Solv.	Name 1	Used	Percent (%)
A	100.0 %	Water V.03	Yes	95.00
B	100.0 %	Acetonitrile V.03	Yes	5.00
C			No	
D			No	

Timetable

Time (min)	A (%)	B (%)	C (%)	D (%)	Flow (mL/min)	Pressure (bar)
12.00	0.00	100.00	0.00	0.00	---	---
15.00	0.00	100.00	0.00	0.00	---	---
18.00	95.00	5.00	0.00	0.00	---	---

Stoptime

Stoptime Mode: Time set
Stoptime: 18.00 min