




Instrument Name LCMS01 Sample Name LB-CVT-01-27

Injection Acq Method QC 5 to 95 MeCN MW 100-800 ESI+-.amx
Name


MWD1B,Sig=220,4 Ref=400,50 PMP1D,Solvent Ratio B



Signal: MWD1B,Sig=220,4 Ref=400,50






RT [min]	Type	Width [min]	Area	Height	Area%	Name
5.227	BV	0.2306	1344.0017	328.6371	28.3120	
5.357	VB	0.1388	85.4062	20.3971	1.7991	
5.597	BV	0.1701	3039.0185	874.0641	64.0183	
5.679	VB	0.1591	268.9695	94.9439	5.6660	
6.169	BB	0.1192	9.7119	3.2360	0.2046	
		Sum	4747.1078			

MWD1D,Sig=280,4 Ref=400,100 PMP1D,Solvent Ratio B



Signal: MWD1D,Sig=280,4 Ref=400,100

RT [min]	Type	Width [min]	Area	Height	Area%	Name
5.228	BB	0.4044	472.8131	113.0042	52.3714	
5.597	BV	0.3222	424.8006	113.9257	47.0533	
5.688	VV b	0.0600	5.1935	2.9311	0.5753	
		Sum	902.8072			





Acquisition Method: QC 5 to 95 MeCN MW 100-800 ESI+-.amx
Path: D:\CDSPProjects\CURVETX\Results\LCMS01-2021-08-17 17-53-58+01-00.rsl

Quat. Pump Method

Flow: 0.500 mL/min
Use Solvent Types: Yes
Low Pressure Limit: 0.00 bar
High Pressure Limit: 600.00 bar
Max. Flow Ramp Up: 100.000 mL/min²
Primary Channel: Automatic
Max. Flow Ramp Down: 100.000 mL/min²
Module Display Name: Quat. Pump
Module Type: G7104C
Order: 1

Solvent Composition

Channel	Ch. 1 Solv.	Name 1	Used	Percent (%)
A	100.0 %	Water V.03	Yes	95.00
B	100.0 %	Acetonitrile V.03	Yes	5.00
C			No	
D			No	

Timetable

Time (min)	A (%)	B (%)	C (%)	D (%)	Flow (mL/min)	Pressure (bar)
12.00	0.00	100.00	0.00	0.00	---	---
15.00	0.00	100.00	0.00	0.00	---	---
18.00	95.00	5.00	0.00	0.00	---	---

Stoptime

Stoptime Mode: Time set
Stoptime: 18.00 min