

Time (mins)	Flow Rate (ml/min)	% Eluent A	% Eluent B
0	0.5	100	0
10	0.5	100	0
70	0.5	60	40
80	0.5	0	100
90	0.5	0	100
93	0.5	100	0
105	0.5	100	0

Table-1 HPLC Running Gradient